

Capello Patient Positioner

Designed by William Capello, MD

Provides stable positioning of a patient during hip procedures

Board is available in a one-piece or two-piece design.

Optional two-piece board construction allows for easier use and storage.

All gel pads, pegs and peg height extensions can be used with existing peg boards.

Optional two-piece board design for easy handling
(Also available in a one-piece design)



PRODUCT NO'S:	
4090	[Set with 2-Piece Board]
4095	[Set with 1-Piece Board]
Replacement Parts:	
4090-PB	[2-Piece Positioning Board]
4095-PB	[1-Piece Positioning Board]
4090-06	[6" (15,2 cm) Radiolucent Peg] Four included in set; one with this product number
4090-08	[9" (22,9 cm) Radiolucent Peg] Four included in set; one with this product number
4090-SC	[Stabilizing Clamp] Two included in set; one with this product number
4090-01	[Large Gel Pad]
9120	[Table Clamp] Two included in set; one with this product number
Optional Parts:	
4090-02	[Peg Gel Pad]
4090-EXT	[Peg Extension – 4" (10,2 cm)]
4090-EXT6	[Peg Extension – 6" (15,2 cm)]
4090-EXT8	[Peg Extension – 8" (20,3 cm)]

Each set includes:
Board, Gel Pad, (4) 6" Radiolucent Pegs,
(4) 9" Radiolucent Pegs, (2) Stabilizing Clamps,
(2) Table Clamps.



6" (15,2 cm) &
9" (22,9 cm)
Radiolucent Pegs

Gel Pad

Stabilizing
Clamp

Table
Clamp

Board

Board Dimensions: 47" x 18.75" (120 cm x 47,6 cm)

– Over –

ISO 13485:2016

FREE TRIAL ON MOST INSTRUMENTS

INNOMED



103 Estus Drive, Savannah, GA 31404
www.innomed.net info@innomed.net

912.236.0000 Phone
912.236.7766 Fax


Innomed-Europe Tel. +41 41 740 67 74
Fax +41 41 740 67 71

1.800.548.2362

© 2021 Innomed, Inc.


Capello Patient Positioner

Set-up guide for patient positioning in total hip surgery





Step 1:

The positioner is placed on the surgical table and secured with the two stabilizing clamps provided – one at the head of the table, and the other at the foot of the positioner on the opposite side of the table. The clamps are secured and locked to the table rails using the two table clamps.



Step 2:

The gel pad is placed on the positioner. A sheet (not shown) can then be placed over the padding to help prevent shifting as the patient is transferred to the table.




Step 3:

Radiolucent pegs should be positioned as follows:

- 1) A short peg placed in the sacral area. A short peg is used here instead of a long peg in order to maximize the operative area.
- 2) A short peg placed in front of the pubis. Care should be taken to avoid neurovascular structure compression in both femoral triangles.
- 3) A long peg placed in the posterior thoracic region below the scapula.
- 4) A long peg placed directly below the pectoral muscle, keeping in mind the need for adequate ventilation of this area.

Note: More radiolucent pegs, as well as the use of the peg extenders, may be required for larger patients.



CAUTION: The board should not be used as a transport device.

– Over –