

STERILIZATION PROCEDURE

Fromm Femur & Tibia Triangle

Product No: 2760

Made of aluminum.

Please use the following guidelines when sterilizing this product:

Sterilizer Type:	Pre-vacuum Steam Sterilizer
Minimum Temperature:	132 degrees C / 270 degrees F
Minimum Cycle Time:	4 minutes
Minimum Dry Time:	30 minutes

Can be pre-vacuum steam sterilized **unwrapped and unassembled** with straps and silicone pad not attached to the triangle.

The silicone pad may be cleaned with a liquid enzymatic detergent, such as Enzol. Please use our standard cleaning procedure for the triangles and pads. We recommend replacing used straps.

Automated Cleaning Cycle

- 1) Disassemble the device.
- 2) Load the instruments into the washer so that the design features are exposed to cleaning.
- 3) Devices capable of holding liquid should be loaded such that the design feature can drain.

The following guidelines are recommended for cleaning. Other automatic cleaning solutions that can be used are Neodisher®, Mediclean Forte®, Thermostept®.

Phase	Time (MM:SS)	Temp. (°C)	Detergent
Prewash	2:00	Cold Water	
Wash	3:00	60+/-5°C	Enzol
Rinse	0:15	60+/-5°C	
Final Rinse	1:00	80+/-5°C-DIW	
Dry Time	6:00	≥ 80°C	

Revised 12/14/2017

FREE TRIAL ON MOST INSTRUMENTS

© 2017 Innomed, Inc.

INNOMED 



103 Estus Drive, Savannah, GA 31404
www.innomed.net info@innomed.net

912.236.0000 Phone
912.236.7766 Fax

1.800.548.2362