

Incavo Tibial Component Revision Osteotomes

Designed by Stephen J. Incavo, MD

*Designed to help break the posterior cement-bone interface
when removing a cemented tibial TKA component*

Also used to help break the posterior implant-bone interface when
removing a cementless tibial TKA component.



PRODUCT NO'S:	
3621-00	[Complete Set]
Set Includes:	
3621-01	[Standard Osteotome] Blade Length: 10mm Blade Width: 1/2" (12.7mm) Blade Offset: 3/4" (19.1mm) Overall Length: 8.5"
3621-02	[Medium Osteotome] Blade Length: 14mm Blade Width: 1/2" (12.7mm) Blade Offset: 3/4" (19.1mm) Overall Length: 8.5"
3621-03	[Deep Osteotome] Blade Length: 18mm Blade Width: 1/2" (12.7mm) Blade Offset: 3/4" (19.1mm) Overall Length: 8.5"
3040	[Slap Hammer]
1015	[Sterilization Case]



FREE TRIAL ON MOST INSTRUMENTS

© 2012 Innomed, Inc.



103 Estus Drive, Savannah, GA 31404
www.innomed.net info@innomed.net

912.236.0000 Phone
912.236.7766 Fax

Innomed-Europe Tel. +41 41 740 67 73
Fax +41 41 740 67 74

1.800.548.2362